
Integrated Electric Intelligent Ultraviolet Germicidal Lamp System UV lamp remote control Instruction Manual

1. Operation

- 1) Device will be in self-checking and initialize state when first time operating. Please do not move and run the device in 1 minute
- 2) Indicator:
Green light(blue-tooth working indicator) : flash light while waiting for onnection;
consecutive light after connecting.
Red light(control device working indicator) : flash light while in normal state;
consecutive light when UV is on.
- 3) Voice report:
Avoid false entering : Do not get into during disinfection.
Expired time : The life of germicidal lamp tube is expired, please replace it.
Damage reminder: The germicidal lamp is damaged, please replace it.

2. APP control and setting


A. Obtain the latest APP program and install on the Android Phone.

Scan the QR code, download "UV lamp remote control XXX.apk"




B. Please click on “approve”when first time using App.

C. Please clink on UV pattern”Sterilizing Lamp”to enter system on the App page.



D. Enter UV light interface,then it will automatically search the surrounded UV system bluetooth device and choose “UV System” .



E. Input password

Default password : 1234

Change password : type 4 numbers for password

Reset password : type super password to restore default password

(super password can be obtained from supplier)

Please enter password to connect to the Bluetooth

Password:

☐ Remember the password

[Change Password](#)

[Reset Password](#)

Enter

Annotations:

- Type the password
- Choose saving password
- Change password
- Reset password
- Confirm then enter control interface

F. Change and reset password

Please type in new password and old password to change password

Old Password: _____

New Password: _____

New Password: _____

Change

Cancel

Please enter the password of Administrator to reset the password

Password: _____

Reset Password


Cancel


G. UV light control and setting:

<

Germicidal Lamp

UV System





Change name.Input method "Enter" save.

The two icons are gray, indicating that the UV lamp and lampshade are in standby and not working.


Click when the icon is gray,
Turn on the UV lamp


Click when the icon is gray,
Open the lampshade

<

Germicidal Lamp

UV System





2 icons in color,Indicates that the UV lamp and lampshade are in working condition.

Click when the icon is colored,
Turn off the UV lamp

Click when the icon is colored,
Close the lampshade

- 4 -

Version:1.1

PS:

- ◆ The lampshade is closed,Click the gray lamp icon,The lamp cover will be opened automatically before the UV lamp is lit.
- ◆ The UV lamp is in working state,Click the colored switch icon,Will turn off the UV lamp first,Then close the lampshade,The machine enters the standby state.